



Essco Calibration Laboratory
 Division of Walsh Engineering Services, Inc.
 27 Industrial Ave, Unit #9, Chelmsford, MA 01824-3618
 Tel: 800-325-2201 978-250-0880 www.esscolab.com
Certificate of Calibration



Issue Date: 07/25/2024

Certificate #: 3172733

Page 1 of 4

CUSTOMER / LOCATION MICROTECHNOLOGIES, INC. @ 128 GARDEN ST. FARMINGTON, CT 06032 PURCHASE ORDER: 2910	EQUIPMENT INFORMATION MANUFACTURER: TEMPERATURE GUARD MODEL #: VM605E SERIAL #: 0717244427 CONTROL #: 1229271 TYPE: TEMPERATURE DATA LOGGER
---	---

PERFORMED: IN LAB AS FOUND: IN TOLERANCE AS LEFT: IN TOLERANCE	Remarks:
--	----------

ENV: TEMP: 20.66°C REL HUMIDITY: 63.79 %RH BAR PRESSURE: 1011.68 hPa	CALIBRATION DATE: 07/25/2024	CALIBRATION DUE: 07/25/2026
--	------------------------------	-----------------------------

Method	Description	Report No.	Last Cal. Date	Cal. Due Date
ECP NO. 1.3.4 Rev: 3 12/7/2020	RTD THERMOMETERS/READOUTS	3021099	01/25/2024	01/25/2025
Standard E5194	FLUKE 5560A MULTIFUNCTION CALIBRATOR			

The ESSCO Quality System is accredited to ISO/IEC 17025:2017

The results above relate only to the item(s) calibrated. This report shall not be used to claim product certification, approval, or endorsement by NVLAP, NIST, or any agency of the United States Federal Government. Any number of factors may cause a nonconformance (i.e. out of tolerance) between calibrations. Expanded Measurement uncertainties were calculated per ISO/IEC Guide 98-3:2008 Guide to the Expression of Uncertainty in Measurement with approximately a 95% confidence level and a coverage factor of k=2. Statements of conformity are made based on the Simple Acceptance binary decision rules as defined in ILAC G8:2019 with measurement uncertainty not taken into account. It is the responsibility of the user to consider the measurement uncertainty when determining compliance to their own processes. The measurement results are traceable to the International System of Units (SI Units) through a National Metrology Institute (NMI), such as NIST, a competent laboratory (i.e. ISO/IEC 17025), or competent producer (i.e. ISO 17034). This calibration was performed in compliance with the ESSCO Quality System manual, ECL1 Rev 50, dated 27 June 2024, ISO/IEC 17025:2017, ISO 9001:2015, ANSI/NC SL Z540-1-1994 part 1, ISO 10012:2003, ISO 13485:2016, and when required contractually, 10 CFR 21 and 10 CFR 50 App. B.

This certificate shall not be reproduced, except in full, without the written approval of Essco.

Adam Vachon

**Adam Vachon
Metrologist**

Philip Thai

**PHILIP THAI
Releasing Authority**

Test #	Function Tested	Nominal	Limits	Unit of Measure	As Found	As Left	Uncertainty
1	INPUT 1, 1000 Ohm RTD	-70.0	-69.0 -71.0	°F	-69.7	-69.7	0.13°F
2	INPUT 1, 1000 Ohm RTD	-15.0	-14.0 -16.0	°F	-14.6	-14.6	0.13°F
3	INPUT 1, 1000 Ohm RTD	36.0	37.0 35.0	°F	36.2	36.2	0.14°F
4	INPUT 1, 1000 Ohm RTD	46.0	47.0 45.0	°F	46.2	46.2	0.14°F
5	INPUT 1, 1000 Ohm RTD	78.0	79.0 77.0	°F	77.8	77.8	0.14°F
6	INPUT 2, 1000 Ohm RTD	-70.0	-69.0 -71.0	°F	-69.8	-69.8	0.13°F
7	INPUT 2, 1000 Ohm RTD	-15.0	-14.0 -16.0	°F	-14.7	-14.7	0.13°F
8	INPUT 2, 1000 Ohm RTD	36.0	37.0 35.0	°F	36.2	36.2	0.14°F
9	INPUT 2, 1000 Ohm RTD	46.0	47.0 45.0	°F	46.2	46.2	0.14°F
10	INPUT 2, 1000 Ohm RTD	78.0	79.0 77.0	°F	78.0	78.0	0.14°F
11	INPUT 3, 1000 Ohm RTD	-70.0	-69.0 -71.0	°F	-69.8	-69.8	0.13°F
12	INPUT 3, 1000 Ohm RTD	-15.0	-14.0 -16.0	°F	-14.6	-14.6	0.13°F
13	INPUT 3, 1000 Ohm RTD	36.0	37.0 35.0	°F	36.2	36.2	0.14°F
14	INPUT 3, 1000 Ohm RTD	46.0	47.0 45.0	°F	46.6	46.6	0.14°F
15	INPUT 3, 1000 Ohm RTD	78.0	79.0 77.0	°F	77.6	77.6	0.14°F
16	INPUT 4, 1000 Ohm RTD	-70.0	-69.0 -71.0	°F	-69.9	-69.9	0.13°F
17	INPUT 4, 1000 Ohm RTD	-15.0	-14.0 -16.0	°F	-14.7	-14.7	0.13°F
18	INPUT 4, 1000 Ohm RTD	36.0	37.0 35.0	°F	36.2	36.2	0.14°F
19	INPUT 4, 1000 Ohm RTD	46.0	47.0 45.0	°F	46.2	46.2	0.14°F
20	INPUT 4, 1000 Ohm RTD	78.0	79.0 77.0	°F	77.7	77.7	0.14°F

Test #	Function Tested	Nominal	Limits	Unit of Measure	As Found	As Left	Uncertainty
21	INPUT 5, 1000 Ohm RTD	-70.0	-69.0 -71.0	°F	-69.8	-69.8	0.13°F
22	INPUT 5, 1000 Ohm RTD	-15.0	-14.0 -16.0	°F	-14.7	-14.7	0.13°F
23	INPUT 5, 1000 Ohm RTD	36.0	37.0 35.0	°F	36.1	36.1	0.14°F
24	INPUT 5, 1000 Ohm RTD	46.0	47.0 45.0	°F	46.1	46.1	0.14°F
25	INPUT 5, 1000 Ohm RTD	78.0	79.0 77.0	°F	77.5	77.5	0.14°F
26	INPUT 6, 1000 Ohm RTD	-70.0	-69.0 -71.0	°F	-69.8	-69.8	0.13°F
27	INPUT 6, 1000 Ohm RTD	-15.0	-14.0 -16.0	°F	-14.7	-14.7	0.13°F
28	INPUT 6, 1000 Ohm RTD	36.0	37.0 35.0	°F	36.1	36.1	0.14°F
29	INPUT 6, 1000 Ohm RTD	46.0	47.0 45.0	°F	46.2	46.2	0.14°F
30	INPUT 6, 1000 Ohm RTD	78.0	79.0 77.0	°F	77.6	77.6	0.14°F
31	INPUT 7, 1000 Ohm RTD	-70.0	-69.0 -71.0	°F	-69.6	-69.6	0.13°F
32	INPUT 7, 1000 Ohm RTD	-15.0	-14.0 -16.0	°F	-14.5	-14.5	0.13°F
33	INPUT 7, 1000 Ohm RTD	36.0	37.0 35.0	°F	36.3	36.3	0.14°F
34	INPUT 7, 1000 Ohm RTD	46.0	47.0 45.0	°F	46.2	46.2	0.14°F
35	INPUT 7, 1000 Ohm RTD	78.0	79.0 77.0	°F	77.6	77.6	0.14°F
36	INPUT 8, 1000 Ohm RTD	-70.0	-69.0 -71.0	°F	-69.9	-69.9	0.13°F
37	INPUT 8, 1000 Ohm RTD	-15.0	-14.0 -16.0	°F	-14.6	-14.6	0.13°F
38	INPUT 8, 1000 Ohm RTD	36.0	37.0 35.0	°F	36.2	36.2	0.14°F
39	INPUT 8, 1000 Ohm RTD	46.0	47.0 45.0	°F	46.3	46.3	0.14°F
40	INPUT 8, 1000 Ohm RTD	78.0	79.0 77.0	°F	77.6	77.6	0.14°F

Test #	Function Tested	Nominal	Limits	Unit of Measure	As Found	As Left	Uncertainty
--------	-----------------	---------	--------	-----------------	----------	---------	-------------

End of Data

Note: A = Adjusted F = Failed L = Limited N/A = Outside Scope of Accreditation