



Temperature Guard NIST Calibration Certificate

Microtechnologies, Inc.
128 Garden St.
Farmington, CT 06032

Calibration traceable to the National Institute of Standards and Technology has been accomplished on the instrument listed below by comparison with standards maintained by Microtechnologies Inc. The accuracy and stability of all standards maintained by Microtechnologies, Inc. are traceable to the national standards maintained by the National Institute of Standards and Technology.

Calibration has been performed by a trained technician using automatic calibration equipment and instruments which are maintained and calibrated in accordance with Microtechnologies' Quality Manual BQ-303

Calibration Results **PASSED**

Channel

Cal Points °F	1	2	3	4	5	6	7	8
-95.0	-95.0	-95.2	-94.8	-94.9	-95.1	-94.8	-94.8	-94.8
78.0	78.0	78.0	77.9	77.9	78.0	77.9	77.9	78.0
32.0	32.2	32.2	32.2	32.3	32.3	32.3	32.3	32.2

Powered by Temperature Guard

Test Results (+- 1 Degree)

Serial Number: 0729244446

Model: VM605E

MAC: 00:80:a3:7a:a0:f5

Calibration Date: 7/30/2024

Due Date: 7/30/2026

Certified By: *JW*

Traceability:

Fluke DMM, Model 8050A, Serial Number: 4030038, Calibration Date 5/30/2024, Due Date 5/30/2026
Calibrated by Essco Calibration Laboratory, 27 Industrial Avenue, Chelmsford, MA 01824. (NVLAP LAB CODE: 200972-0)

Essco Calibration Laboratory has been audited and found in compliance with ISO/IEC 17025:2017